

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 07 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130132</u> ✓
Name <u>Matthew E. Armstrong</u>		ID# <u>5906</u> Date <u>5 May 2009</u> ✓
A	Agency <u>Seward PD</u>	Phone # <u>(907) 224-3338</u>
Instrument Location <u>410 Adams St Seward Community Jail, Seward AK</u> ✓		
B		
Alco S/N <u>X301605</u> ✓		Target Value <u>.082</u> ✓ High Pressure <u>900</u>
Alco Test Values		
<u>.075</u> ✓		<u>.076</u> ✓
1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco
Signature <u>Matthew E. Armstrong</u> ✓		<u>CMS</u> <u>5/18/09</u>
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.


Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May, 2009.

Carolyn M. Noland  
Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130132

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATA MASTER cdm 130132

MAY 05, 2006

OPERATOR'S NAME:

OPERATOR'S NUMBER: 0006

SUBJECT'S NAME:

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

SUBJECT'S FIRST NAME: WT

SUBJECT'S LAST NAME: WT

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATA MASTER cdm 130132

MAY 05, 2006

TIME 14:00

DIAGNOSTIC CHECK

COMPUTER: OKAY

PERIPHERAL: OKAY

SOFTWARE DATE: 05/29/06

WATERS: OKAY

WATER CHANGER: 48c

WATER USED: 44c

BAROMETER: 29.54 in

FLOW DETECTOR: OKAY

PUMP: OKAY

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

11/11/2005 14:00:00 DATA MASTER cdm 130132

11/11/2005 14:00:00 DATA MASTER cdm 130132

11/11/2005 14:00:00 DATA MASTER cdm 130132